



4.5.2 – MAINTENANCE LOG BOOK

[Signature]



PRINCIPAL
K.M.C.T. AYURVEDA
MEDICAL COLLEGE



KMCT

AYURVEDA MEDICAL COLLEGE

Approved by NCISM and Affiliated to Kerala University of Health Sciences.

TO WHOM SO EVER IT MAY CONCERN

This is to certify that the information in the attached documents is verified by me and is true to the best of my knowledge.



S. John
Principal

PRINCIPAL
K.M.C.T. AYURVEDA
MEDICAL COLLEGE

no	Equipment	complaint	department
1	microscope	not working	brauyader
2	Autoclave	not working	nidancam department
3	Boyles apparatus	not working	salayatent
4	IV stand	damaged	shalyatent
5	Barron pilos glen	damaged.	shalyatent
6	Electrical suction machine.	not working	shalyatent
7	vertical BP instrument	not working	shalyatent

Date	Action taken	attended by	verified by
24/5/2019	checked & change magnifying lens	Rajesh	de
14/9/2019	checked the machine & changed.	Rajesh	de
15/10/19	checked the machine & changed.	Joyel	
17/11/19	checked & replaced.	Joyel	de
18/12/19	checked & replaced.	Joyel	
22/12/19	checked & replaced.	Joyel	de
28/12/19	checked & replaced.	Joyel	de

no	Equipment	Complaint	Department
8	oil immersion microscope	not working	Shanivark
9	sigmoidoscope suction machine	root not working	Shalyastan
10	tonsilular suction	not working	Shalyastan
11	Thermometer	not working	kayachikitsa
12	weighing machine	not working	kayachikitsa
13	stethoscope	not working	Shalyastan
14	BP apparatus	not working	Shalyastan
15	Examination table	damaged.	pravartan

Date	Action Taken	Attended by	Verified by
2/11/20	checked & corrected	Rajesh	Q
3/2/20	checked & replaced	Rajesh	Q
4/4/20	checked & replaced	Rajesh	Q
4/6/20	checked & replaced	Rajesh	Q
5/7/20	checked & replaced	Rajesh	Q
12/18/20	checked & replaced	Joyel	Q
2/20/20	checked & replaced	Joyel	Q
24/10/20	Repaired	Joyel	Q

no	Equipment	complaint	department
16.	Calorimeter	not working	Kryosher
17	Refrigerator	not working	Kryosher
18	centrifuge	not working.	Kryosher
19.	Mixer	not working	Reactor
20	electronic balance	not working.	Reactor
21	hot plate	not working.	Reactor
22	weighing machine	not working.	Rajewyannan
24	Bp apparatus	not working	panchakri
25	microscope	not working.	Drayegur

Date	Action Taken	attended by	verified by
20/11/20	checked & repaired.	Joyel	le
20/12/20	checked & repaired	Joyel	le
27/12/20	checked & replaced.	Rosh	le
3/1/21.	checked & repaired	Rosh	le
4/2/21.	checked & replaced	Joyel	le
4/3/21.	checked & replaced	Joyel	le
14/5/21	checked repaired	Rosh	le
21/7/21.	checked & replaced	Rosh	le
14/9/21.	checked & repaired	Joyel	le

Sno	Equipment	Complaint	Department
26.	weighing machine	not working	kaumara bhrtya.
27.	thermo meter	not working.	panchaka
28	vacuum Extractor	not working	practitioner
29	sterilized	not working	pracutina
30	Radient warmer	not working.	pracutina
31	pulse oxy meter	not working.	pracutna
32	Binocular microscope.	not working.	Rogawyna
33	Hot air oven	not working.	Rogawyna
34	dissection table	damaged.	Anatomy

	Date	Action Taken	attended by	verified by
1	21/11/21.	checked & replaced.	Rajesh	le.
2	12/12/21.	checked & replaced.	Arun	le.
3	4/12/21.	checked & repaired	Rajesh	le.
4	5/12/21.	checked & repaired	Joyel	le.
5	4/4/22.	checked & repaired	Joyel	le.
6	5/6/22.	checked & replaced	Arun	le.
7	17/18/22.	checked & repaired	Joyel	le.
8	27/10/22.	checked & repaired	Joyel	le.
9	30/11/22.	checked & replaced	Joyel	le.

SNo	Equipment	Competent	Not working	Reported
35.	Hot plate		not working	Rasika
36.	Autoclave		not working	Rajawyn
37.	Urinometer		not working	Rajawyn
38	Centrifuge graduate machine		not working	Rajawyn
39.	Bunsen burner		not working	Rajawyn
40	Nebulizer		not working	Prachi
41.	Suction machine		not working	Shayam
42	vertical BP instrument		not working	Shayam
43	IV stand		damaged	Shayam

No.	Date	Action taken	checked by	verified by
1.	4/2/23.	checked & replaced	Rajesh	K.
2.	5/6/23.	checked & replaced	Arun	K.
3.	12/7/23.	checked & replaced	Arun	K.
4.	11/9/23.	checked & replaced	Rajesh	K.
5.	21/10/23.	checked & replaced	Joyel	K.
6.	14/11/23	checked & replaced	Joyel	K.
7.	27/11/23.	checked & replaced	Rajesh	K.
8.	11/12/24.	checked & replaced	Rajesh	K.
9.	12/12/24.	checked & replaced	Joyel	K.

sl no	Equipment	complaint	Department
44.	Autoclave.	Not working	Widanaar departmr
45	oil immersion microscope	Not working.	Shrivard
46	micro	Not working.	Rakractra
47	Refrigerator	not working.	Kryosare

Date	Action taken	attended by	venue/where
20/3/24.	checked & repaired.	Joyel.	Ge.
21/5/24.	checked & replaced	Joyel.	Ge.
13/7/24.	checked & repaired	Joyel.	Ge.
20/7/24.	checked & repaired	Joyel.	Ge.